

Certificate of Analysis

Product name: Eurotrol HemoTrol

Reference number: **REF** 171.002.002

Lot number: **LOT** 23166

Expiry date:  2024-02-29

Manufacturing date:  2022-08-01

Assigned Values:

HemoCue Hb systems	Hemoglobin		
	Unit	Mean	Range
HemoCue Hb 201+ system	g/dL	12.0	11.4 - 12.6
	g/L	120	114 - 126
	mmol/L	7.4	7.0 - 7.8

The mean assay value provided on the enclosed assay sheet is derived from replicate analyses on factory calibrated HemoCue systems. The values are only valid on the HemoCue Hb 201 system. Results obtained should fall within the expected range specified for the control.

Bioburden: A bioburden test is performed and no micro-organisms were found.

Homogeneity: The product successfully passed homogeneity testing.

Leakage test: A leakage test is performed and there was no product loss.

Filling volume: 1 ± 0.05 mL

Sample selection: All used vials for listed tests are randomly selected from the produced batch.

Date: 2022-08-11

This document was created digitally and is valid without signature.

AN01119A01